


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-2W-C-M-05-P*

S.N.: *2346*

Date: *APRIL 25 1985*



- a) Output signal amplitude:
0 TO +10 AMPS
- b) Pulse width:
0 TO 50 NSEC
- c) Rise time:
≤ 1 NSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*≈ 7 AMPS/DIV
10 NSEC/DIV
10 KHz*

*120/240V
50-60 Hz*