


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-V-HV3-C-PN-M-05*

S.N.: *2347*

Date: *APRIL 30 85*



- a) Output signal amplitude:  
*0 TO  $\pm 100$  VOLTS*
- b) Pulse width:  
*0 TO 100 NSEC*
- c) Rise time:  
 *$\leq 200$  PSEC*
- d) Fall time:  
 *$\leq 3$  NSEC*
- e) PRF:  
*0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*100 VOLTS/DIV*

*5 NSEC/DIV*

*5 KHZ*

*120/240V*

*50-60 HZ*

*[Signature]*