


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-N-PS*

S.N.: *2354*

Date: *APRIL 30 85*



- a) Output signal amplitude:
0 TO -15 VOLTS.
- b) Pulse width:
0 TO 2 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240V

50-60 Hz

*10 VOLTS/DIV
1.0 NSEC/DIV
5 MHz.*