


PULSE GENERATOR

PERFORMANCE CHECK

Model: *APMA-2-N-PS*

S.N.: *2355*

Date: *APRIL 30 85*



a) Output signal amplitude:

*0 TO 15 VOLTS*

b) Pulse width:

*0 TO 2 NSEC.*

c) Rise time:

*≤ 100 PSEC.*

d) Fall time:

*≤ 135 PSEC.*

e) PRF:

*0 TO 25 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*120/240 V*

*50-60 Hz.*

*10 VOLTS/DIV*

*1.0 NSEC/DIV*

*5 MHz*

*[Signature]*