


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APM-2-N*

S.N.: *2364*

Date: *APRIL 30 85*



a) Output signal amplitude:

*0 TO 15V*

b) Pulse width:

*0 TO 2 NSEC*

c) Rise time:

*≤ 100 PSEC*

d) Fall time:

*≤ 135 PSEC*

e) PRF:

*0 TO 25 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+24 ± 1.0 VOLT*

*600 mA (MAX)*

*10 VOLTS/DIV*

*1.0 NSEC/DIV*

*5 MHz*

*[Signature]*