


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV4-2-e-2N

S.N.: 2377

Date: APRIL 18 85



- a) Output signal amplitude:  
0 TO  $\pm 125$  V<sub>PP</sub>
- b) Pulse width:  
 $\leq 2$  nsec.
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 100 kHz
- f) Jitter, stability:  
OK
- g) Prime power:

100 V<sub>PP</sub> / DIV  
2 nsec / DIV  
20 kHz

120/240 V  
50-60 Hz  
A