


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANM-1-C-D-PN*  
S.N.: *2380*  
Date: *MAY 14 '85*



- a) Output signal amplitude:  
*0 TO ± 5 VOLTS*
- b) Pulse width:  
*0 TO 6 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*5 VOLTS/PIV*  
*1.0 NSEC/PIV*  
*5.0 MHz*

*120/240 V*  
*50-60 Hz*  
*AA*