


PULSE GENERATOR
PERFORMANCE CHECK

Model: 1720-1-P

S.N.: 2386

Date: MAY 7 85



- a) Output signal amplitude:
25 TO 100 AMPS
- b) Pulse width:
≈ 25 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:
OK

≈ 60 AMPS
10 NSEC/DIV
1 KHz

+15 ± 1.0 VOLT
200 mA (MAX)
[Signature]