


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APP-AV-1-N*

S.N.: *2403*

Date: *MAY 27 1985*



- a) Output signal amplitude:
0 TO -10 V ...
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 VOLT
250 mA (max)

1.0 VOLT/DIV
1.0 NSEC/DIV
500 KHz.