


PULSE GENERATOR
PERFORMANCE CHECK

Model: ANI-NRL1-PS

S.N.: 2405

Date: JUNE 17 85



- b) Output signal amplitude:
 $\geq +60$ VOLTS TO
- b) Pulse width: 50 Ω .
4 NSEC.
- c) Rise time:
 ≤ 50 PSEC
- d) Fall time:
 < 2 NSEC.
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:

32 VOLTS/DIV
2 NSEC/DIV
50 KHz

120/240V
50-60 Hz

- h) PROPAGATION DELAY:
 ≤ 2 NSEC.
- i) INPUT TOLERANCE
 $\geq +2.2$ VOLTS TO 50 Ω
10W 10 NSEC TO 1 μ SEC.

[Signature]