


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVMM-1-P-M-C-D

S.N.: 2410

Date: MAY 10 85



- a) Output signal amplitude:
0 TO +5 V.
- b) Pulse width:
0 TO 6 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC.
- e) PRF:
0 TO 25 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV.
1.0 NSEC/DIV.
5 MHz.

120/240 V
50-60 Hz