

PULSE GENERATOR  
PERFORMANCE CHECK

Model:


AVL-2-C-PN-T

S.N.:

2414


Date:

APRIL 25 1985



(A)

≈ 320 VOLT/DIV  
10 NS/DIV  
(MIN TR)



(B)

AS (A) BUT  
MAX TR

a) Output signal amplitude:

0 TO ± 350 V

b) Pulse width:

0 TO 100 NS

c) Rise time:

≤ 1.5 NS

d) Fall time:

→ 10 NS

≤ 3 NS

e) PRF:

→ 10 NS

0 TO 5 KHZ

f) Jitter, stability:

O.K.

g) Prime power:

120/240V

50-60 Hz