

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AVMN-1-P

S.N.:

2415

Date:

APRIL 17 85



a) Output signal amplitude:

0 TO 15 V. RMS

b) Pulse width:

0 TO 6 NSEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 200 PSEC

e) PRF:

0 TO 50 MHz

f) Jitter, stability:

OK

g) Prime power:

$\pm 24 \pm 1.0$ V

500 mA (MAX)

5 VOLTS/DIV

100 NSEC/DIV

40 MHz.