

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVDMN-1-N

S.N.:

2413

Date:

APRIL 17 85



a) Output signal amplitude:

0 to 5 VOLTS

b) Pulse width:

0 to 6 USEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 200 PSEC

e) PRF:

0 to 50 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 ± 1.0 VOLTS

500 MA (MAX)

*[Signature]*

5.0 VOLTS/DIV

1.0 USEC/DIV

40 MHz