


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMMN-1-N*  
S.N.: *2419*  
Date: *APRIL 4 85*



- a) Output signal amplitude:  
*0 TO 5 VOLTS*
- b) Pulse width:  
*0 TO 6 NSEC.*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 200 PSEC*
- e) PRF:  
*0 TO 50 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*5 VOLTS/DIV*  
*1.0 NSEC/DIV*  
*5.0 MHz*

*+24 ± 1.0 VOLTS*  
*500 mA (MAX)*