


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-5B-C-M-P

S.N.: 2422

Date: APRIL 25 1985



- a) Output signal amplitude:
0 TO +48 AMPS.
- b) Pulse width: (TO 0.75 μ s)
25 TO 500 μ SEC.
- c) Rise time:
 \leq 10 NSEC
- d) Fall time:
 \leq 10 NSEC.
- e) PRF:
0 TO 1 KHz.
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240V
50-60 Hz

[Signature]

10 VOLTS/DIV
(TO 0.7 μ s)
100 NSEC/DIV
1.0 KHz