



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR-3-PS-PN*  
 S.N.: *2431*  
 Date: *JUNE 20 85*



P<sub>out</sub> *50 VOLTS/DIV*  
*5 USEC/DIV*

- a) Output signal amplitude:  
*0 TO ± 200 VOLTS TO 50 Hz*
- b) Pulse width:  
*0.1 TO 100 USEC*  
*(EQUALS INPUT PW)*
- c) Rise time:  
*≤ 10 NSEC*
- d) Fall time:  
*≤ 10 NSEC*
- e) PRF:  
*MIN PW & 0 TO 5 KHz*  
*MAX PW: 0 TO 50 Hz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*



N<sub>out</sub> *50 VOLTS/DIV*  
*5 USEC/DIV*

*[Signature]*