


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-PM2-P*

S.N.: *2444*

Date: *JUNE 28 85*



- a) Output signal amplitude:
> +60 V
- b) Pulse width:
≤ 2 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*32 VOLTS/DIV
5.0 NSEC/DIV
100 KHZ*

*+135 ± 5 VOLTS
5 mA (MAX)*