


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-AU-1-PS-PN-M*


S.N.: *2462*

Date: *JUNE 14 '85*



- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
0 TO 4 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV.
100 NSEC/DIV
500 KHz.*

*120/240 V
50-60 Hz
*