


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-P*  
S.N.: *2468*  
Date: *MAY 21 85*



- a) Output signal amplitude:  
*0 TO +40 VOLTS*
- b) Pulse width:  
*2 TO 100 NSEC*
- c) Rise time:  
*≤ 350 PSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*SEE PLS 1, 2*  
*OK*
- g) Prime power:

*32 VOLTS/DIV*  
*2 NSEC/DIV*  
*40.0 KHz*

*+15 ± 1.0 VOLT*  
*125 mA (max)*