


PULSE GENERATOR
PERFORMANCE CHECK

Model: 19VMN-1-P

S.N.: 2472

Date: MAY 16 85



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0 TO 6 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+24 ± 1.0 VOLTS
500 mA (MAX)

5 VOLTS/DIV
5 NSEC/DIV
40 MHz