


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-P-C*

S.N.: *2479*

Date: *JUL 7 11 85*



- a) Output signal amplitude:
25 TO 100 AMPS
- b) Pulse width:
≈ 25 NSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHz.
- f) Jitter, stability:
OK
- g) Prime power:

*≈ 60 AMPS/DIV
10 NSEC/DIV
100 KHz.*

*120/240 V,
50 - 60 Hz*

[Signature]