


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM R -2 - P - LA1*

S.N.: *2480*

Date: *MAY 13 85*



- a) Output signal amplitude:
+ 20 VOLT TO 50 Ω
- b) Pulse width:
TO 40 NSEC
- c) Rise time:
 \leq 300 PSEC
- d) Fall time:
 \leq 300 PSEC
- e) PRF:
0 TO 10 MHZ
- f) Jitter, stability:
OK
- g) Prime power:
 - a) *+ 36 \pm 1.0 VOLT 200mA MAX*
 - b) *-10 \pm 3VOLT 50mA MAX*

10 VOLT/DIV
5 NSEC/DIV
10 MHZ
R_L = 50 Ω