


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PR-1-PW-PS-P*

S.N.: *2494*

Date: *Jul 7 10 '85*



- a) Output signal amplitude:  
*0 TO +200 V*
- b) Pulse width:  
*0.7 TO 5 USEC*
- c) Rise time:  
*≤ 10 NSEC*
- d) Fall time:  
*≤ 10 NSEC*
- e) PRF:  
*0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50 VOLT/DIV*  
*0.5 USEC/DIV*  
*1.0 KHz*  
*R<sub>c</sub> = 50 Ω*

*120/240V*  
*50-60 Hz*

*[Signature]*