


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-1-D-P*

S.N.: *2499*

Date: *JULY 12 85*



- d) Output signal amplitude:  
*0 TO +5 VOLTS*
- e) Pulse width:  
*0 TO 6 NSER*
- f) Rise time:  
*≤ 100 PSEC*
- g) Fall time:  
*≤ 135 PSEC*
- h) PRF:  
*0 TO 25 MHz*
- i) Jitter, stability:  
*OK*
- j) Prime power:  
*+24 ± 1.0 VOLTS*  
*500 mA (MAX)*

*5 VOLTS/DIV.*  
*1.0 NSER/DIV*  
*5.0 MHz.*