


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVO-5B-P-C

S.N.: 2503

Date: JULY 19 '85



- a) Output signal amplitude: 0 TO +48 AMPS
- b) Pulse width: 20 TO 500 NSEC
- c) Rise time:  $\leq$  10 NSEC
- d) Fall time:  $\leq$  10 NSEC
- e) PRF: 0 TO 1 KHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V  
50-60 Hz.

100 NSEC/DIV  
10 VOLTS/DIV TO  
 $R_L = 0.72 \Omega$   
1 KHz.

120/240V  
50-60 Hz.