


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-2-C-N-PN-M*

S.N.: *2506*

Date: *AUG 8 65*



- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
5 TO 50 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
2 NSEC/DIV
50 KHz

120/240V
50-60 Hz.

[Signature]