



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVN-1-C-P*  
S.N.: *2515*  
Date: *JULY 25 85*



Ⓐ *5 nSEC/DIV*  
*5 VOLTS/DIV*



Ⓑ *AS Ⓐ*  
*BUT 1 nSEC/DIV.*

- a) Output signal amplitude:  
*0 TO +5 VOLTS*
- b) Pulse width:  
*0.2 TO 1.0 nSEC.*
- c) Rise time:  
*≤ 100 pSEC.*
- d) Fall time:  
*≤ 100 pSEC*
- e) PRF:  
*50 TO 250 AMHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*[Signature]*