


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVL-TN-1*

S.N.: *2524*

Date: *JULY 22 85*



- a) Output signal amplitude:
-230 VOLTS & +30 VOLTS
- b) Pulse width:
590 100 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 3 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*≈ 100 VOLTS/DIV
10 NS/DIV*

*120/240 V
50-60 Hz*

h) PROPAGATION DELAY

≤ 100 NS