


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMN-1-P*

S.N.: *2530*

Date: *JUNE 19 '85*



- a) Output signal amplitude:  
*0 TO +5 VOLTS*
- b) Pulse width:  
*0.2 TO 6.0 NSEC.*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 200 PSEC*
- e) PRF:  
*0 TO 50 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*5 VOLTS/DIV*  
*5 NSEC/DIV*  
*240 MHz*

*+24 ± 1.0 VOLT*  
*500 mA (MAX)*