


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMN.-1-N*

S.N.: *2532*

Date: *JUNE 19 '85*



- a) Output signal amplitude:
0 TO -5VOLT
- b) Pulse width:
0.2 TO 6 NSER
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLT/DIV
5 NSER/DIV
≈ 40 MHz

+24 ±1.0 VOLT
500 mA (MAX)
[Signature]