


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-AN-1-CP*

S.N.: *2533*

Date: *JULY 23 85*



- a) Output signal amplitude:  
*0 TO +100 V.*
- b) Pulse width:  
*0 TO 100 NSEC*
- c) Rise time:  
*≤ 1 NSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 TO 5 KHZ.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240V*  
*30-60 KHZ.*

*100 VOLTS/DIV*  
*10 NSEC/DIV*  
*5 KHZ.*  
*R<sub>L</sub> = 50Ω*