


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-AV-1-e-p*

S.N.: *2534*

Date: *JUL 24 85*



- a) Output signal amplitude:
0 to +100V
- b) Pulse width:
0 to 100 nsec
- c) Rise time:
≤ 1 nsec
- d) Fall time:
≤ 2 nsec
- e) PRF:
0 to 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240V
50-60 Hz

100 VOLTS/DIV
10 nsec/DIV
5 kHz
R_L = 50 Ω

[Signature]