


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANMR-2-C-TRF-OT-P*


S.N.: *2537*

Date: *OCT 2 85*



- a) Output signal amplitude:
0 TO +20 VOLTS
- b) Pulse width:
20 TO 200 NSEC
- c) Rise time:
0.3 AND 3 NSEC
- d) Fall time:
0.3 AND 3 NSEC
- e) PRF:
0 TO 5 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240V
50-60 Hz

A) TR, TF 176 H
5 NSEC/DIV
10 VOLTS/DIV
1 MHz



B) AS A BUT
TR, TF LOW.