


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AV14-MPT-P

S.N.: 2539

Date: JULY 17 85



- a) Output signal amplitude:  
71 + 30 0.483
- b) Pulse width:  
≤ 1.0 nsec.
- c) Rise time:  
N/A
- d) Fall time:  
N/A
- e) PRF:  
0 TO 1 MHz.
- f) Jitter, stability:  
OK
- g) Prime power:

32 VOLT/DIV  
1.0 nsec/DIV  
500 KHz.

+15 ± 10 V  
150 mA (MAX)  
[Signature]