


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APR-1-DW-PS-N*

S.N.: *2546*

Date: *JULY 10 85*



- a) Output signal amplitude:  
*0 TO -200V.*
- b) Pulse width:  
*0.1 to 5 USEC*
- c) Rise time:  
*≤ 10 NSEC*
- d) Fall time:  
*≤ 10 NSEC*
- e) PRF:  
*0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*

*50 VANTS/DIV*  
*1.0 USEC/DIV*  
*1 KHz*  
*R<sub>L</sub> = 50 Ω*

*120/240V*  
*50-60 Hz*  
*[Signature]*