

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AVM-1-N

S.N.:

2555

Date:

JUN 12 85



a) Output signal amplitude:

0 TO -5 VOLTS

b) Pulse width:

0 TO 1 NSEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 ± 1.0 VOLT

500 mA (MAX)

5 VOLTS/DIV

1.0 NSEC/DIV

5.0 MHz