


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-A1-PW-PS-PN-CA-EW*

S.N.: *2558*

Date: *MAR 25 86*




- a) Output signal amplitude:
0 TO 200 VOLTS
- b) Pulse width:
50 NSEC TO 1.0 msec
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC

- e) PRF:
MAX PW: 0 TO 5 KHz
MIN PW: 0 TO 100 KHz
- f) Jitter, stability: *(0.5% max DUTY CYCLE)*
OK

- g) Prime power:
120/240 V
50-60 Hz

[Signature]

P_{OUT} *50 VOLTS/DIV*
100 NSEC/DIV
R_L = 50 Ω



N_{OUT} *50 VOLTS/DIV*
100 NSEC/DIV
R_L = 50 Ω