


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AN-1-C-DN*

S.N.: *2566*

Date: *JULY 4 85*



- b) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width:
0 TO 4 μ SEC
- e) Rise time:
 ≤ 100 PSEC
- b) Fall time:
 ≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- b) Jitter, stability:
OK
- b) Prime power:

10 VOLTS DIV
100 μ SEC DIV
1.0 MHz

120/240 V
50-60 Hz
[Signature]