


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMA N-1-P*

S.N.: *2575*

Date: *JUN 24 85*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0 to 6 μ SEC
- c) Rise time:
 ≤ 100 PSEC
- d) Fall time:
 ≤ 200 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*5.0 dB/DIV
5 nSEC/DIV
(≈ 40 MHz)*

*+24 \pm 1.0 VOLT
500 mA (MAX)*