


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ACORN-1-N*
S.N.: *2578*
Date: *AUG 25 85*



- a) Output signal amplitude:
0 TO -5 V
- b) Pulse width:
0 TO 6 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 50 KMHZ
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
5 NSEC/DIV
≈ 40 KMHZ

+24 ± 1.0 V
400 mA (MAX)
[Signature]