


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-N-1-P*

S.N.: *2580*

Date: *AUG 25 85*



- d) Output signal amplitude:  
*0 TO +5 V. OUT*
- e) Pulse width:  
*0 TO 6 NSEC*
- f) Rise time:  
*≤ 100 PSEC*
- g) Fall time:  
*≤ 200 PSEC*
- h) PRF:  
*0 TO 50 MHz*
- i) Jitter, stability:  
*OK*
- j) Prime power:

*5 VOLTS/DIV*  
*5 NSEC/DIV*  
*≈ 40 MHz.*

*+24 ± 1.0 V*  
*400 mA*