


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-3-PS-M-ED*

S.N.: *2599*

Date: *SEPT 26 85*



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
2 TO 15 NSEC.
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 600 PSEC
- e) PRF:
0 TO 25 MHZ.
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
1 NSEC/DIV
5 MHZ

120/240V
50-60 HZ