


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-S-1-C7

S.N.: 2603

Date: SEPT 9 85



- a) Output signal amplitude:
0 TO +10V.
- b) Pulse width:
≤ 135 PSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
500 PSEC/DIV
100 KHz

120/240V
50-60 Hz
[Signature]