


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-3A-P*

S.N.: *2613*

Date: *OCT 15 '85*



- a) Output signal amplitude:
+25 to +75 Amps
- b) Pulse width:
2 to 5 NSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLT

*100 VOLTS/DIV
($R_L = 2 \Omega$)*

2 NSEC/DIV

200 mA (MAX)