


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVP-AV-15-C-PN
-05-M-CA-EW

S.N.: 2614

Date: OCT 25 85



- a) Output signal amplitude:
0 TO ± 5 VOLTS.
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
 ≤ 60 PSEC
- d) Fall time:
 ≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
1.0 NSEC/DIV
1.0 MHz

120/240V
50-60 Hz