


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2-IFT1-N-05-M*

S.N.: *2622*

Date: *AUG 21 85*



- a) Output signal amplitude:
0 TO -350 V.
- b) Pulse width:
0 TO 100 NSEC.
- c) Rise time:
≤ 1.5 NSEC
- d) Fall time:
≤ 3 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK.
- g) Prime power:

*≈ 320 VOLTS/DIV.
10 NSEC/DIV
1.0 KHz.*

*+15 ± 1.0 VOLT
500 mA MAX.*