


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-2-1TT1-N-05-M*

S.N.: *2623*

Date: *APR 21 '85*



- a) Output signal amplitude:  
*0 TO - 350 V ...*
- b) Pulse width:  
*0 TO 100 NSEC*
- c) Rise time:  
*≤ 1.5 NSEC*
- d) Fall time:  
*≤ 3 NSEC*
- e) PRF:  
*0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*≈ 320 VOLTS/DIV  
10 NSEC/DIV  
1.0 KHZ*

*+15 ± 1.0 VOLT  
500 MA (MAX)*

*[Signature]*