


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMM-3-P*
S.N.: *2625*
Date: *SEPT 25 85*



- a) Output signal amplitude:
0 TO +15V ± 4%
- b) Pulse width:
2 TO 15 NSEC
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 600 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
2.0 NSEC/DIV
10 MHz

+24 TO +28 V
400 mA (MAX)

[Signature]