


PULSE GENERATOR
PERFORMANCE CHECK


Model: AP-14-15-C-N

S.N.: 2629

Date: JULY 30 85



Ⓐ 5 VOLTS/DIV
1.0 NSEC/DIV
1.0 MHz



Ⓑ AS Ⓐ BUT
100 PSEC/DIV

- a) Output signal amplitude:
0 TO -5 V
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Signature]